

Average temperature  
difference: logger–WRS (°C)

Year	No. of tests	Average temperature difference: logger–WRS (°C)		
		00:00– 24:00	10:00– 16:00	00:00– 06:00
2004	216	−0.02	0.13	−0.09
2005	67	−0.01	0.10	−0.08
2006	6	−0.13	−0.04	−0.20
2007	38	−0.05	0.16	−0.16
2008	48	−0.20	−0.25	−0.16
2009	73	−0.22	−0.26	−0.20
2010	140	−0.21	−0.23	−0.19
2011	83	−0.23	−0.36	−0.20