

Data for Modern Soil Chronometry using Fallout Radionuclides

Joshua D. Landis¹

¹Dept. Earth Sciences, Dartmouth College, 19 Fayerweather Hill Road, Hanover, NH USA 03755

Correspondence to: joshua.d.landis@dartmouth.edu

Abstract

We describe environmental gamma spectrometry data for >700 soil samples collected from >35 high-resolution quantitative soil profiles spanning global sites. The data are collected for the purpose of modern soil chronometry based on fallout radionuclides (FRNs) ⁷Be and ²¹⁰Pb, using the Linked Radionuclide Accumulation model (LRC). Cumulative gamma counting time for samples in the database exceeds 6.5 years. This is a living database to be augmented as data become available and corrected with improvements in data reduction or identification of errors. Versions and changes will be indexed. Special attention is paid to measurement uncertainties in the dataset, as well as how atmospheric or excess ²¹⁰Pb is defined in both geochemical and mathematical terms for use in the LRC model. Basic familiarity with gamma spectrometry and radionuclide decay chains is assumed. The data set can be accessed at <https://doi.org.10.17632/cfxkpn6hj9.1> (Landis, 2025).

Short Summary

Understanding rates of environmental change is critical to human and ecological health but is difficult when the processes are too slow or too small to observe directly. To overcome this limitation, we can use natural radioactive elements as virtual 'clocks' to measure change. Here we describe a large number of measurements that have been used to develop soils as clocks or chronometers of change to atmospheric carbon and mercury (Hg) cycles.

1. Introduction

We describe high-precision measurements of environmental radionuclides collected for the purpose of developing modern soil chronometry. Numerous problems such as atmospheric pollutant deposition and carbon sequestration require knowledge of the rates of soil processes that influence metals and carbon through time, but an appropriate chronometry system spanning years to centuries of environmental change has not been available until recent efforts to validate fallout radionuclide (FRN) chronometry (Landis, 2023; Landis et al., 2016, 2025; Landis, Obrist, et al., 2024; Landis, Taylor, et al., 2024). Historically, the impediment to adoption of FRNs in soil systems has been the availability of sufficient instrumental resources to provide long counting times that are necessary for measurement of short-lived ⁷Be, in datasets large enough to develop appropriate age models (Kaste et al., 2011). Here, we report the fallout radionuclides ⁷Be and ²¹⁰Pb, together with bomb pulse ²⁴¹Am and ¹³⁷Cs, and ancillary radiogenic isotopes including ²²⁶Ra, ²²⁸Th, ²²⁸Ra, and others, that are used to construct and verify the Linked Radionuclide Accumulation model (LRC). These radionuclides are all measured concurrently by gamma spectrometry in the Fallout Radionuclide Analytics facility (FRNA) at Dartmouth College. The data are accessible at <https://doi.org.10.17632/cfxkpn6hj9.1>; Landis, 2025).

LRC is an empirical age model that dates soil organic and mineral matter based on a mass balance comparison between the soil accumulative ⁷Be:²¹⁰Pb ratio (R_{acc}) and the ⁷Be:²¹⁰Pb atmospheric flux ratio

(R_0)(Landis et al., 2016). The data presented currently include more than 700 samples from over 35 high-resolution quantitative soil pits collected using field methods adapted for this purpose, and more are being added each month (Fig. 1). The method is confirmed by accurately dating the bomb-pulse peak in ^{241}Am (or ^{137}Cs where appropriate) which is known to have occurred in 1963-64 (Health and Safety Laboratory, 1977). The method is further corroborated by both novel ^{228}Th : ^{228}Ra chronometry and ^{14}C approaches and yields uncertainties on the order of 10% up to 100 years and 20% up to 200 years (Landis, 2023; Landis et al., 2025).

The majority of sampled soils are what we refer to as 'reference', selected to minimize the potential for historical disturbance through development, land use, natural erosion or tree throw over the past minimum of 150 years. Additional sites may include experimental designs examining hillslope and alluvial or fluvial processes and should not be taken as representative of reference conditions, i.e. with an expectation that the nuclear bomb-pulse can be accurately dated. Implementation of the LRC model is described elsewhere (Landis et al., 2016). Additional details and challenges of environmental gamma spectrometry (Cutshall et al., 1983; Landis et al., 2012; Murray et al., 1987, 2018) or theories of related FRN age models can also be found elsewhere (Arias-Ortiz et al., 2018; Barsanti et al., 2020; Sanchez-Cabeza & Ruiz-Fernández, 2012). Any brief discussion of these topics here is intended only to contextualize the data and thereby provide an entry point for non-specialist readers. A basic understanding of gamma spectrometry and radionuclide decay chains is assumed.

Finally, data validation is discussed with respect to both analytical problems such as instrument calibration, as well as geochemical problems, namely the susceptibility of Rn to diffuse out of samples in the laboratory. A priority of this manuscript is to demonstrate that routine environmental gamma spectrometry is capable percent-level precision that is expected from other capital analytical platforms such as inductively-coupled plasma optical or mass spectrometry used for elemental and isotopic analyses (ICPOES, ICPMS).




Figure 1. Site map for soils collected for modern soil chronometry (Google Earth Pro, data SI, NOAA, U.S. Navy, NGA, GEBCO. Image IBCAO, Landsat, Copernicus. Altitude 5181 miles).

80

2. Methods

2.1. Instrumentation

85 Data are collected with six Mirion (formerly Canberra) Broad Energy intrinsic germanium detectors (BEGe) with carbon fiber endcaps and ultra-low background cryostats which are cooled with liquid nitrogen. The detectors are shielded with 4 inches of ultra-low background lead. Data acquisition is performed with Mirion DSA-1000 or Lynx digital signal processors running with Genie 2K software v.3. The specific BEGe detector design is important because it provides excellent efficiency at low gamma energies for ^{210}Pb acquisition without sacrificing efficiencies at high energies. Carbon fiber provides good transmission of the low energy ^{210}Pb emission which is otherwise attenuated by an aluminum endcap. 90 Samples are typically run using four-day count times. Of two detector models, BEGe 3830 (38 cm² detector active area, 30 mm thickness) and larger BEGe 5030 (50 cm²), the latter saves one day of counting and accomplishes in three days what the 3830 accomplish in four.

2.2. Sample preparation and geometry

95 Foliage, litter, and soil samples are oven dried at 60 °C. Bulk soils are sieved or hand-picked to remove clasts and debris >2mm in diameter. Fine roots >1 mm diameter or >2 cm in length are removed by hand. Foliage and leaf litter (Oi and Oe horizon) are typically milled to <1 mm using a Wiley mill with stainless steel blades. Soil samples are loosely disaggregated as necessary by agate mortar and pestle and packed into 110 cm³ polyethylene petri dishes with snap closures. The petris are sealed with common electrical tape and rested for >2 weeks to allow for ^{222}Rn ingrowth, although we will show that 100 this approach does not quantitatively retain Rn (see also (Murray et al., 1987, 2018)). Our handling of the problem of Rn emanation from both laboratory samples and soils *in situ* is discussed below.

2.3. Calibration and Sample Self-Absorption

105 Gamma detector calibration is based on geometric efficiency, which is the proportion of photons emitted by the sample that are recorded by the detector, absent sample self-absorption of photons which is dependent on its specific elemental composition. Net efficiency includes self-absorption of photons by the sample. We measure net efficiency directly by using a series of soils/sediment doped 1% by weight with either uranium ore BL5 or thorium ore OKA2 which are available from Natural Resources Canada 110 Certified Reference Material Project. Geometric efficiency for the standards is then calculated as the quotient of net efficiency and the standard self-absorption coefficient at each photopeak energy in the (described below). The efficiencies are representative of the detector-sample geometric relationship and applicable to all sample unknown measurements. Efficiencies for photopeaks not represented in the standards are estimated by regression from adjacent photopeaks in the U or Th standards (e.g., ^{7}Be , ^{241}Am , or ^{137}Cs ; these are not known to have coincidence summing effects). Reported uncertainties on 115 the standard activities is 0.2% for U and 1.0% for Th, and secular equilibrium is assumed for both.

Net efficiency for individual samples is calculated as the product of geometric efficiency determined from standards and the sample self-absorption at each energy of interest (Landis et al., 2012). To do this, 120 self-absorption for each sample is measured using a planar uranium ore source (same diameter as sample) placed directly on top of the sample. This is performed as a second data acquisition following the sample 'long count' which does not use point source. Self-absorption at all energies is calculated using the method of (Cutshall et al., 1983), or in the case of non-uranium series radionuclides, is estimated by linear regression from the U-series photopeaks.

125

Calibration uncertainty estimated from geometric efficiencies of 6 standards of bulk density ranging from 0.5 g cm⁻³ to 200 g cm⁻³ is less than 2% relative standard deviation for each of our detectors. Uncertainty

of sample self-absorption derived from peak integration (below) contributes <1%. Thus, our net efficiency calibration uncertainties for most radionuclides are <2% at the one-sigma level.

2.4. Peak integration

A semi-automated peak integration routine is implemented in Excel as described by (Landis et al., 2012). For each detector, peak shape parameters are measured across energies either within the collected sample spectrum using peaks >10,000 counts or using prior collection of a high-activity point source spectrum. A peak shape model is then used to estimate the partial area of sample peaks that are integrated digitally by summation. Partial peak integration reduces measurement uncertainty by avoiding peak tails which contribute disproportionately to integration error.

2.5. Spectral Background Correction

All radionuclide peaks are corrected for background contributions which are measured with composite background spectra totaling hundreds of days of count time. Background for each peak may variously include ambient Rn in the laboratory, artifacts of the gamma spectrum, impurities in the detector housing and shielding, and ambient construction materials in the laboratory facility.

2.6. Data Quality and Verification

Environmental gamma spectrometry is capable of producing percent-level precision on par with other capital analytical platforms such as ICP-MS. Precision in gamma spectrometry estimates of radionuclide activity is easily documented because many radionuclides emit photons at multiple energies. These typically have variable coincidence summing effects that require correction at each energy, as well as sample self-absorption factors that also require correction at each energy. Concordance among estimates from multiple photopeaks spanning the energy spectrum thus provides a robust estimate of precision in the gamma measurements.




Figure 2. Correlations of multiple photopeaks for (a) ^{222}Rn , (b) ^{228}Th , (c) ^{228}Ra . Lines show linear best fits.

For quality control measures we recommend reporting multi-line means and uncertainties for three radionuclides: ^{222}Rn measured at ^{214}Pb (295, 352 keV) and ^{214}Bi (609 keV); ^{228}Ra measured with up to four emission lines from ^{228}Ac (209 keV, 338 keV, 463 keV, 911 keV); and ^{228}Th measured by ^{212}Pb (238 keV), ^{208}Tl (538 keV), and possibly ^{224}Ra (241 keV) (Fig. 2).

The coefficient of variance (CV) or relative standard deviation (RSD) for multiple emission peaks for each of ^{222}Rn , ^{228}Th , and ^{228}Ra provides an assessment of measurement accuracy and precision in the gamma spectrum (Fig. 3). These are typically in the 1-5% range for typical crustal activity concentrations of 20 Bq kg^{-1} . ^{222}Rn is important since it is often used as an indirect measure of ^{226}Ra with integration uncertainties 5-10 times lower (Fig. 3a), but with the important caveat that the actual values of the two measures differ substantially due to geochemical disequilibrium caused by Rn leakage (Sect. 3.5).




Figure 3. Average and variance-weighted average activities for radionuclides measured at multiple energies in the gamma spectrum, including (a) ^{222}Rn , with comparison to direct measure of ^{226}Ra ; (b) ^{228}Th ; (c) ^{228}Ra . ^{226}Ra measurement is possible at a single emission line which is compromised by low photon yield and photopeak convolution with ^{235}U , resulting in uncertainties 5-10 times higher than possible with e.g. ^{222}Rn daughter radionuclides.

2.6. Peak interference deconvolution

Multiple important radionuclides in the gamma spectrum suffer significant peak-peak overlap that requires deconvolution. Most notably are ^7Be which has interference from both ^{228}Ac (478 keV) and ^{214}Pb (280 keV) (Landis et al., 2012), and ^{226}Ra which has interference from ^{235}U at 185 keV and which approximately doubles its analytical uncertainty (Zhang et al., 2009). Our ^7Be deconvolution regime is verified elsewhere using repeat measures of a single natural sample since the known half-life of ^7Be can be exploited to predict its change in concentration over monthly timescales (Landis et al., 2012). The ^{226}Ra deconvolution regime can be tested against certified reference materials; however, the reported uncertainties for ^{226}Ra in available standards typically exceeds 10% and are therefore of limited use in verifying robust measure of this radionuclide (more below in Sect. 8).




Figure 4. (a) Measurement of ^{224}Ra is possible in the gamma spectrum but requires deconvolution of its 241 keV emission from ^{214}Pb at 242 keV. (b) Accuracy of the deconvolution procedure can be tested against ^{212}Pb (via ^{220}Rn) measured at 238 keV.

Deconvolution can also be tested using the measure of ^{224}Ra (half-life 3.6 days, 241 keV) as a proxy for its parent radionuclide ^{228}Th because of its photopeak overlap with ^{214}Pb at 242 keV. The peak deconvolution procedure is evaluated with the expectation that radioactive equilibrium is maintained, i.e., with a slope of 1 in correlation between the parent and daughter (Fig. 4). Here the slope = 0.985 $\pm$ 0.006 which shows a systematic error of just 1.5 $\pm$ 0.6%.

3. Radionuclide measurements

3.1. ^7Be and problem of detection. The defining limitation in LRC development is the problem of ^7Be detection. The combination of low natural abundance, low photon yield (10.5%), and short half-life (54 days) requires extraordinary efforts in both the logistics of sample collection and availability of instrument time to make these measurements possible. Samples should be measured within 1 month of collection and must be counted for 4 days or more per sample to define the ^7Be soil depth profile (Fig. 5). Following these protocols ^7Be may be detected to depths of 10 or more centimeters in humid or moist climates, which has important implications both for how soils are perceived to function and how age models are applied to soils (Landis et al., 2016).




Figure 5. Elapsed measurement of the ^7Be gamma photoregion over 1, 2, 3, and 4-day count times. The measured sample is soil from 5 cm depth, and ^7Be activity is quantified as $1.8 \pm 1.0 \text{ Bq kg}^{-1}$ after 4 days. The modeled ^7Be peak is shown in black, U-series peaks in blue, and Th-series peaks in red. MDL is shown with dashed line.

3.2. ^{241}Am and problem of detection. Demonstrating accuracy of LRC requires concordance with independent chronometers, and bomb-pulse ^{241}Am is the prime candidate due to its long half-life, limited solubility or bioavailability (in contrast to ^{137}Cs), and potential for its concurrent measurement in the gamma spectrum. The bomb-pulse has a well-known history of activity in the atmosphere (Appleby et al., 1991; Health and Safety Laboratory, 1977). Abundance of ^{241}Am is low, however, rarely exceeding 1 Bq kg^{-1} in temperate soils, and is altogether undetectable in equatorial soils where fallout deposition was 10-times lower than mid-latitudes (Chaboche et al., 2021). Alternate measurement of Pu isotopes by ICPMS or AMS is becoming routine and may offer an alternative where ^{241}Am cannot be measured (e.g. Alewell et al., 2017)).

3.3. ^{137}Cs and problem of geochemical mobility. In contrast to ^{241}Am , ^{137}Cs is easily measured in the gamma spectrum. However, it can be subject to biological cycling as a K analog (Kaste et al., 2021). As a result, the shape of Am and Cs profiles can diverge remarkably (Fig. 6). ^{241}Am is always the preferred measure of bomb pulse. Still, where ^{241}Am and ^{137}Cs profiles appear consistent, ^{137}Cs may provide insight into shape of the ^{241}Am profile when latter is detection limited.

3.4. Total ^{210}Pb . Total ^{210}Pb is the sum of both atmospheric (excess) and geogenic (supported) sources and is easily measured in gamma spectrum with BEGe detectors. The strongest source of error is the self-absorption correction since this can vary by a factor of 2-3 through the soil profile. However, the low photon yield (4%) can make measurement of supported ^{210}Pb challenging in soils with typical uranium abundance. The four-day count time is thus important for providing adequate precision for defining $^{210}\text{Pb}_{\text{ex}}$ at the tail of its depth distribution (where $^{210}\text{Pb}_{\text{ex}}$ is calculated as $^{210}\text{Pb}_{\text{total}}$ minus ^{222}Rn ; see below).




Figure 6. Depth profiles of ^{241}Am and ^{137}Cs in diverse soils. In all cases the cumulative ^{241}Am date is consistent with bomb-pulse peak of 1963-64. Assigned calendar years are as follows: BrT05 =1970.1, DoF01 =1965.3, HuB01 =1964.4, CaPr01 =1964.2, HoF01 =1958.1. In some cases, ^{137}Cs follows ^{241}Am closely, in others (especially spodosol) their divergence is pronounced and ^{137}Cs follows a "leaky cycle" whereby some portion may be strongly retained near the soil surface while any escaping this surface peak is rapidly flushed through the soil profile below.

3.5. ^{226}Ra and problem of radon emanation. The measure of ^{226}Ra is important, in principle, for defining the activity of ^{210}Pb that is supported in a soil by secular equilibrium with U-bearing mineral content (discussed in detail in Sect. 4). Secular equilibrium requires that radionuclides in a decay chain have equal activities, and we should observe this in the chain $^{226}\text{Ra} > ^{222}\text{Rn} > ^{214}\text{Bi} > ^{214}\text{Pb} > ^{210}\text{Pb}$ provided that gaseous Rn can be retained within the sample. Instead, however, we observe that ^{214}Pb and ^{214}Bi underestimate ^{226}Ra even in samples that are sealed and rested for the required ingrowth period for Rn to achieve equilibrium (5 half-lives or about 20 days) (Fig. 7). Disequilibrium between ^{222}Rn and ^{226}Ra is difficult to assess in a single sample, however, because of high uncertainties on the ^{226}Ra measure due in part to significant overlap from the primary ^{235}U emission at 185.7 keV (Zhang et al., 2009).

In this dataset we quantify disequilibrium between ^{222}Rn and ^{226}Ra as $f\text{Rn}$, or fraction Rn, which is simply the ratio of the two measures. (We acknowledge that the problem is somewhat more complicated because the effective volume of Rn in the detector shield will be greater than volume of the sample itself due to its diffusion from the sample container into surrounding air). Among all soil pits the mean measured $f\text{Rn} = 79 \pm 1\%$ ($\pm\text{RSE}$, $n=30$) (Fig. 7). Importantly, the variance of $f\text{Rn}$ within a single pit is much smaller than among pits [$p < 0.0001$], which means that $f\text{Rn}$ is a property of the geochemistry of individual soils (with respect to hosting of Ra in primary vs. secondary minerals, grain size, susceptibility to weathering, geological age, etc.).




Figure 7. Rn retention by soils, (a) fraction ^{222}Rn retained per sample in laboratory, calculated as $^{222}\text{Rn}/^{226}\text{Ra}$, The mean rate of ^{222}Rn disequilibrium in this dataset on a per-sample basis is $77 \pm 15\%$ (SD, $n=472$, omitting samples with $^{226}\text{Ra} < 20 \text{ Bq kg}^{-1}$). (b) fraction ^{222}Rn retained per soil pit, (c) laboratory disequilibrium between ^{210}Pb and ^{222}Rn in deep soils, per soil pit, calculated as $\epsilon\text{Pb} = ^{210}\text{Pb}/^{222}\text{Rn}$.

3.6. Experimental confirmation of ^{226}Ra accuracy and ^{222}Rn losses.

Affirming the accuracy of ^{226}Ra measurements is foundational to interpreting ^{210}Pb in the soil profile. Here we confirm the accuracy of direct ^{226}Ra measurements (186 keV) against certified reference materials (CRMs) (Fig. 8), where performance is excellent with standard 4-day count times (Fig. 8a). We note that analytical uncertainties thus achieved are typically far better than those of the certification process itself. This reflects the difficulties in direct measure of ^{226}Ra requiring corrections for self-absorption and deconvolution (Tasker et al., 2019).




Figure 8. (a) Certified reference materials for ^{226}Ra determination, IAEA-385 (Irish sea sediment), IAEA-447 (moss-soil), CCRMP CLV-1 (spruce needle), NIST-4353a (Rocky Flats soil). Eckert and Ziegler single isotope solution (EZ ^{226}Ra) is used as an in-house calibration material. Error bars are 1σ standard deviations for analysis at FRNA, larger error bars show certification uncertainties. (b-e) Distributions of results for IAEA-385 certification measurements (red) and FRNA analysis (green) for individual emission lines of ^{226}Ra and the ^{222}Rn daughters ^{214}Pb and ^{214}Bi .

The IAEA-385 material is the most rigorously-calibrated soil standard available to date and critical to our ability to validate high-quality gamma spectrometry (International Atomic Energy Agency, 2013; Pham et al., 2008). Even so, the original recommended ^{226}Ra activity of $22.7 \pm 0.7 \text{ Bq kg}^{-1}$ (Pham et al., 2005) is 3.6% higher than the eventual certification value of 21.9 Bq kg^{-1} . The lower value derives from the

averaging of direct ^{226}Ra measurements with lower estimates from ^{214}Pb (352 keV) and ^{214}Bi (609 keV) (Fig. 8b-e). The latter should be viewed as suspect since both Rn emanation and coincidence summing losses can bias the ^{222}Rn daughters too low. FRNA measurements appear to show a systemic ^{222}Rn loss of 3.4% as the difference between ^{226}Ra activities (22.7 ± 0.3) and ^{222}Rn (21.98 ± 0.07) [$p=0.0186$].

For sample analysis we sought to confirm the accuracy of our ^{226}Ra measures and therefore that Rn disequilibrium is due to Rn loss and not an analytical artifact (Syam et al., 2020). We performed an experiment with NIST-traceable ^{226}Ra standard solution (Eckert and Ziegler; certified uncertainty <2%), which is important to verify that ^{226}Ra is measured accurately and with high enough activity to measure it with percent-level precision (here approximately 360 Bq kg^{-1}). A second calibrated ^{228}Ra solution was prepared by dissolution of thorium ore OKA-2. Accuracy of both Ra source calibrations was previously confirmed by isotope-dilution thermal ionization mass spectrometry (Landis et al., 2018). Briefly, both the ^{226}Ra and ^{228}Ra standard solutions were co-precipitated in a wet synthesis of MnO_2 from KMnO_4 and MnCl_2 at pH ~ 9 . To replicate a typical soil sample in 110 cm^3 petri geometry, the synthesized Ra- MnO_2 was air dried onto a bed of quartz sand. The Ra-Mn-sand thus produced was packed into our standard polypropylene petri as for a soil sample.

The petri was first closed with electrical tape and measured serially to observe Rn ingrowth under our typical sample preparation. Yields of both ^{226}Ra and ^{228}Ra are within error of 100% of certified or expected values which confirms robust preparation of the test material (Fig. 9). Short-lived ^{220}Rn (aka thoron, half-life =56 seconds) is fully retained by the sample, but grows in following the half-life of intermediary ^{224}Ra (3.6 days) which is expelled from the initial suspension of MnO_2 by alpha recoil.




Figure 9. Experimental confirmation of ^{222}Rn emanation from 110 cm^3 soil laboratory samples, (a) ^{222}Rn and ^{226}Ra source, (b) ^{220}Rn and ^{228}Ra source. Closed symbols show yield of ^{226}Ra or ^{228}Ra from the experiment, open symbols show corresponding Rn yield or equilibrium as the Rn/Ra activity ratio.

For ^{222}Rn we observed an equilibrium f_{Rn} of just 55% after approximately thirty days. The sample was then resealed using hot glue, but this reduced Rn retention, presumably, due to porosity in the glue that visibly forms upon cooling. Finally, the sample was encased in 2 mm of paraffin wax. This final step increased f_{Rn} to 85%, but the impracticality of the method coupled with imperfect Rn retention is a fatal strike against its routine use in our laboratory.

Our subsequent efforts to seal samples under vacuum in aluminized Mylar bags improves Rn retention up to 90-95% for our 110 cm³ petri (Mauring & Gäfvert, 2013), but we have not standardized this approach due to **remaining uncertainty in precise quantification of retention**, the creation of additional laboratory waste and, **in the face of these problems**, the uncertain value of improving the indirect measure of ²²⁶Ra to defining ²¹⁰Pb_{ex} (see below).

4. Defining excess Pb-210 (²¹⁰Pb_{ex}).

FRN age models are based on the atmospheric component of ²¹⁰Pb, which in soils and sediment must be separated mathematically from ²¹⁰Pb that is supported by *in situ* production from U-series decay in soil minerals. By convention atmospheric ²¹⁰Pb is measured as the excess over that supported (²¹⁰Pb_s) by its long-lived radioactive precursor ²²⁶Ra, i.e., ²¹⁰Pb_{ex} = ²¹⁰Pb minus ²²⁶Ra. Due to analytical constraints described above, ²²⁶Ra is routinely measured indirectly using ²²²Rn with the assumption that equilibrium is maintained between the two. We have already shown this is not accomplished in typical sample preparation.

Further, however, is the problem of Rn emanation from soils *in situ*: the very process that produces excess ²¹⁰Pb in the atmosphere yields a ²¹⁰Pb deficit in soil. This means that supported ²¹⁰Pb in soils is likely to be *overestimated* if measured by ²²⁶Ra. This leaves two open questions, (1) what is the correct metric for supported ²¹⁰Pb *in situ*, and (2) how do we measure this in routine data collection in the laboratory? To clarify these problems, we introduce the term Pb enrichment, εPb, which is the ratio of ²¹⁰Pb to ²²²Rn in deep soil samples believed free of atmospheric ²¹⁰Pb. Importantly, for all soil pits we have measured εPb = 102 ± 3% (±RSE, n=30) which means that supported ²¹⁰Pb is typically in equilibrium with measured ²²²Rn. This is best illustrated in the soil depth profile where total ²¹⁰Pb converges asymptotically with ²²²Rn, whereas both are overestimated by ²²⁶Ra (Fig. 10). Interestingly, this implies that Rn loss from the measured sample is comparable to loss in the soil *in situ*, and, as observed above, that Rn emanation from standard laboratory samples relates to a physical property of the soil itself.




Figure 10. Defining excess ²¹⁰Pb in soils. (a) Hubbard Brook Experimental Forest (HuB01), (b) Experimental Lakes Area soil 302-1, (c) Harvard Forest LTER (HaF01). Total ²¹⁰Pb typically converges with ²²²Rn at depth which indicates that ²²²Rn is an appropriate estimate of supported ²¹⁰Pb.

We therefore conclude the following regarding the definition of ²¹⁰Pb_{ex} in soils,

Deleted: both

Deleted: ultimately

Deleted: 9

Deleted: 9

- ²²⁶Ra is not an appropriate direct measure of supported ²¹⁰Pb due to chronic emanation ²²²Rn from the soils *in situ*, which results in large ²¹⁰Pb deficiencies in deep soils which are isolated from deposition of atmospheric ²¹⁰Pb_{ex}.
- in situ* ²²²Rn emanation in the soil profile can be assessed directly by gamma measurements, as an asymptotic difference between ²²⁶Ra and ²¹⁰Pb in deep soils.
- measure of ²²²Rn by daughters ²¹⁴Pb and ²¹⁴Bi provides excellent precision and reproducibility among multiple emission lines, 5-10 times better than possible with direct measure of ²²⁶Ra.
- Rn emanation from the laboratory sample is both pronounced and yet systematic despite sealing of containers and incubating >18 days following standard practice.
- nonetheless, ²²²Rn measured in the sample typically converges with ²¹⁰Pb, or is asymptotic with ²¹⁰Pb, suggesting that ²²²Rn measure in the laboratory sample is the best available option for defining ²¹⁰Pb_{ex} in soil.

For defining ²¹⁰Pb_{ex} we use the following expression, noting that this accommodates cases where εPb is not equal to 1, by assuming a constant emanation rate throughout the soil profile (Landis et al., 2016):

$$^{210}\text{Pb}_{\text{ex}} = ^{210}\text{Pb}_T - ^{222}\text{Rn}(\epsilon\text{Pb}) \quad (1)$$

Propagated uncertainties in ²¹⁰Pb_{ex} when calculated by Eq. 1 are shown in Fig. 11.




Figure 11. Propagated uncertainties for ²¹⁰Pb_{ex} as function of (a) total ²¹⁰Pb_{ex} activity, (b) LRC soil age.

5. Demonstrating LRC concordance with bomb pulse ²⁴¹Am

Demonstrating concordance of LRC ages with independent chronometers is important for verifying both the suitability of soils for chronometry and veracity of data production (Landis, Obrist, et al., 2024). We report ²⁴¹Am activities for this purpose, but it should be noted that peak bomb-pulse deposition is indicated at the layer of maximum ²⁴¹Am density (Bq m⁻² cm⁻¹) rather than activity concentration since the latter is skewed by changes in bulk density especially at the soil surface. ²⁴¹Am density is readily calculated from the data provided. In some cases, the shape of ²⁴¹Am bomb-pulse peak can be grossly distorted due to acceleration and deceleration of colloid migration in the soil profile (Landis, Taylor, et al., 2024), and in these cases the cumulative ²⁴¹Am curve becomes an alternate tool for identifying peak bomb fallout since the history of ²⁴¹Am in the atmosphere is well known (Appleby et al., 1991; Health and Safety Laboratory, 1977).

6. Assessing uncertainties on radionuclide measurements.

Our reported uncertainties for radionuclide activities are internal estimates which are propagated from uncertainties on peak integration, which includes background subtraction and deconvolution. Uncertainties due to efficiency calibration and self-absorption correction are not included because, in part, we have not established the degree to which these may be autocorrelated among photo energies. However, external precision can be estimated conservatively by propagating an additional 2% uncertainty as described in Sect. 2.3.

6.1. Z-scores to assess uncertainty estimates

The accuracy of reported integration uncertainties can be assessed using multiple emission lines for a single radionuclide, since each must conform to the same true mean activity value. We show this by means of Z-score, which is the difference between emission line activities, normalized by the propagated uncertainty for the difference. The standard deviation of Z should be equal to 1 if the estimated integration uncertainties accurately represent the true distribution of random error in the measurements (Fig. 12). We observe standard deviations of 0.8, 1.2, and 1.4 for ^{222}Rn , ^{228}Th , and ^{228}Ra , respectively. These are close to 1 which implies that the integration uncertainties are robust. In the case of ^{228}Th and ^{228}Ra the real uncertainties are somewhat underestimated which means that uncertainties in calibration contribute an additional error, and one that is similar in scale to uncertainties in integration, i.e., if similar errors are propagated the result is a 40% higher than either, $\sqrt{1^2+1^2}=1.4$.




Figure 12. Distribution of Z-scores for multiple emission lines of (a) ^{222}Rn (295, 352 keV); (b) ^{228}Th (238, 583 keV); (c) ^{228}Ra (338, 911 keV). Z expresses the difference between two measures in units of σ and is calculated as the difference between two photopeak measures of activity, divided by propagated uncertainty of the two measures.

On the other hand, our prior assessment suggests that integration uncertainties are overestimated by on the order of 30%, based on the unique experiment of measuring radioactive decay of ^7Be since the true value of the sample is known based on its rate of decay, and estimated uncertainties can be evaluated against the observed dispersion of measurements around this known value (Landis et al., 2012).

Overestimation of integration errors is consistent with our observed $Z < 1$ for the ^{222}Rn lines comparison above. That uncertainties for ^{228}Th and ^{228}Ra are greater than for ^{222}Rn suggests that the interaction of coincidence-summing and self-absorption effects yields higher error in calibration, since these are not handled independently in our calibration routine, i.e., by using a Th point source instead of U for Th-

series photopeaks. Future work is required to determine whether one ^{228}Ra measurement may be preferred over others (e.g., among the ^{228}Ac photopeaks at 209, 338, and 911 keV).

7. Spectrum accuracy evaluated using $^{228}\text{Th}:$ ^{228}Ra

The $^{228}\text{Th}:$ ^{228}Ra ratio is an important metric for evaluating accuracy of gamma measurements since it can generally be assumed to have an equilibrium value of 1. To ensure this is true we omit samples with LRC ages <20 years or ^{228}Ra <10 Bq kg⁻¹ since these may have true $^{228}\text{Th}:$ ^{228}Ra disequilibrium due to variable uptake of Th and Ra during plant growth (Landis, 2023). The mean $^{228}\text{Th}:$ ^{228}Ra ratio for the data set is 1.015 ± 0.038 SD, ± 0.002 SE ($n=390$), showing a minor bias of 1.5%. Importantly, the standard deviation for Z (for the difference between $^{228}\text{Th}:$ ^{228}Ra and 1) = 1.20, which shows that the estimates of uncertainties are reasonable (Fig. 13).




Figure 13. Assessing analytical accuracy using the measured $^{228}\text{Th}:$ ^{228}Ra ratio in soils. (a) measured $^{228}\text{Th}:$ ^{228}Ra ratio for all soils, (b) Z-score for difference between measured ratio and 1.

8. Certified Reference Materials

Availability of a few high-quality certified reference materials provides a means for external validation of environmental gamma data. We recommend IAEA-385 (sediment), IAEA-447 (moss soil), and NRC CLV-1 (spruce twig), since these have rigorous quality control and uncertainties low enough (typically <5%) to allow verification of high-precision methods. Performance of our methods against these standards for radionuclides in addition to ^{226}Ra is reported elsewhere with errors against certified values <5% (Landis et al., 2012, 2016).

9. Detection limits

Detection limit applies to total radioactivity in the detector shield [Bq] which is the product of mass and concentration, which means that it is dependent on sample bulk density since the sample volume is fixed. Detection is also inversely dependent on sample count time following Poisson counting statistics. Detection is also influenced by composition of the sample with respect to other elements and radionuclides because these influence photopeaks backgrounds through Compton scattering. Because detection limits vary based on sample mass, composition, and count time they are not easily compiled and compared. We instead defer to reported uncertainties so that the value or weight of each datum can be independently assessed by users on this basis. Given this caveat, estimated detection limits for four-day count times, 100-gram samples, and approximately crustal-average U and Th content are approximately as follows: ^7Be (0.5 Bq kg⁻¹); ^{210}Pb (1.5 Bq kg⁻¹); ^{137}Cs (0.3 Bq kg⁻¹); ^{241}Am (0.15 Bq kg⁻¹); ^{186}Ra (2.5 Bq kg⁻¹); ^{222}Rn (0.2 Bq kg⁻¹); ^{228}Th (0.2 Bq kg⁻¹); ^{228}Ra (1.0 Bq kg⁻¹);

475 **10. Data versions, omissions, and error codes in the database**

Multiple data reduction generations are represented in the database, with notable differences being the introduction of new photopeaks such as 209, 338, and 583 keV in later versions. Data fields are empty in preceding analyses.

480 In some rare cases, samples that are unable fill a 110 cm³ petri are run instead in 10 cm³ petris, with the lower volume incurring a corresponding loss in data quality due to low total activity and resulting counting statistics. Because data quality for the smaller 10 cm³ petri is poor and export fields are not aligned with data reduction for the standard 110 cm³ petri, these data are not reported in the database.

485 The error #NUM! is a peak integration error typically resulting from error in uncertainty estimation for integration of small peaks that are near detection. This can propagate through deconvolution regimes, i.e. error in uncertainty of ²²⁸Ac integration will produce error in uncertainty for ⁷Be.

11. Reported data fields

490 The reported data are tabular in form. Row one is data identifier 1. Row two is data 2, and row three is units or identifier 3. Subsequent rows list sample data per sample. Reported data include the following:

1. pit code: unique identifier for the soil pit.
2. sample name: unique identifier for the soil pit depth interval.
3. sample description: including horizon and/or depth interval.
4. horizon code: soil taxonomic horizon description.
5. interval mass [kg m⁻²]: total mass of fine soil fraction (<2 mm) for the collected depth interval.
6. soil depth [cm]: starts at soil surface =0 rather than at mineral soil surface.
7. R_{acc} : accumulative ⁷Be:²¹⁰Pb ratio for overlying soil layers used in implementation of LRC model.
8. propagated uncertainty (1-***sigma***) for R_{acc} .
9. relative standard deviation (***RSD***) of R_{acc} .
10. count time [seconds] for collected gamma spectrum.
11. activity concentrations [Bq kg⁻¹] for the following radionuclides, with their propagated internal uncertainties: ²¹⁰Pb (46 keV), ⁷Be (477 keV), ²⁴¹Am (60 keV), ¹³⁷Cs (662 keV), ²³⁴Th (63 keV), ²²⁶Ra (186 keV), ²²²Rn (295, 352, 609 keV), ²²⁸Th (238, 538 keV), ²²⁸Ra (338, 911 keV), ⁴⁰K (1460 keV).

505

12. Data availability statement

The data described in this manuscript are freely available to the public (<https://doi.org/10.17632/cfxkpn6hj9.1>; Landis, 2025).

510 **13. Conclusions**

The central limitation to advancing chronometry of modern terrestrial biogeochemistry, including atmosphere-soil interaction and soil processes, is availability of required radionuclide data. We hope that by providing a large, open-access dataset we can help catalyze interest among new research groups and funding agencies in making new investments towards similar data acquisition efforts.

515

14. Competing Interests

The authors declare that no competing interests influence this work.

520

15. References

- Alewell, C., Pitois, A., Meusbürger, K., Ketterer, M., & Mabit, L. (2017). 239 + 240Pu from “contaminant” to soil erosion tracer: Where do we stand? In *Earth-Science Reviews* (Vol. 172, pp. 107–123). Elsevier B.V. <https://doi.org/10.1016/j.earscirev.2017.07.009>
- Appleby, P. G., Richardson, N., & Nolan, P. J. (1991). 241Am dating of lake sediments. *Hydrobiologia*, 214(1), 35–42. <https://doi.org/10.1007/BF00050929>
- Arias-Ortiz, A., Masqué, P., García-Orellana, J., Serrano, O., Mazarrasa, I., Marbá, N., Lovelock, C. E., Lavery, P. S., & Duarte, C. M. (2018). Reviews and syntheses: 210Pb-derived sediment and carbon accumulation rates in vegetated coastal ecosystems - Setting the record straight. *Biogeosciences*, 15(22), 6791–6818. <https://doi.org/10.5194/bg-15-6791-2018>
- Barsanti, M., García-Tenorio, R., Schirone, A., Rozmaric, M., Ruiz-Fernández, A. C., Sánchez-Cabeza, J. A., Delbono, I., Conte, F., De Oliveira Godoy, J. M., Hejnis, H., Eriksson, M., Hatje, V., Laissaoui, A., Nguyen, H. Q., Okuku, E., Al-Rousan, S. A., Uddin, S., Yli, M. W., & Osvath, I. (2020). Challenges and limitations of the 210Pb sediment dating method: Results from an IAEA modelling interlaboratory comparison exercise. *Quaternary Geochronology*, 59(May). <https://doi.org/10.1016/j.quageo.2020.101093>
- Chaboche, P. A., Saby, N. P. A., Lacey, J. P., Minella, J. P. G., Tiecher, T., Ramon, R., Tassano, M., Cabral, P., Cabrera, M., da Silva, Y. J. A. B., Lefevre, I., & Evrard, O. (2021). Mapping the spatial distribution of global 137Cs fallout in soils of South America as a baseline for Earth Science studies. In *Earth-Science Reviews* (Vol. 214). Elsevier B.V. <https://doi.org/10.1016/j.earscirev.2021.103542>
- Cutshall, N. H., Larsen, I. L., & Olsen, C. R. (1983). Direct analysis of 210Pb in sediment samples: self-absorption corrections. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 206(1–2), 309–312. <http://linkinghub.elsevier.com/retrieve/pii/0167508783912735>
- Health and Safety Laboratory. (1977). *90Sr and 89Sr in monthly deposition at world land sites*.
- International Atomic Energy Agency. (2013). *Reference Sheet for Certified Reference Material IAEA-385 natural and artificial radionuclides in sediment from the Iris Sea*. http://nucleus.iaea.org/rpst/Documents/IAEA_AL_151.pdf
- Kaste, J. M., Bostick, B. C., Heimsath, A. M., Steinnes, E., & Friedland, A. J. (2011). Using atmospheric fallout to date organic horizon layers and quantify metal dynamics during decomposition. *Geochimica et Cosmochimica Acta*, 75(6), 1642–1661. <https://doi.org/10.1016/j.gca.2011.01.011>
- Kaste, J. M., Volante, P., & Elmore, A. J. (2021). Bomb 137Cs in modern honey reveals a regional soil control on pollutant cycling by plants. *Nature Communications*, 12(1). <https://doi.org/10.1038/s41467-021-22081-8>
- Landis, J. D. (2023). Age-Dating of Foliage and Soil Organic Matter: Aligning 228Th:228Ra and 7Be:210Pb Radionuclide Chronometers over Annual to Decadal Time Scales. *Environmental Science & Technology*, 57(40), 15047–15054. <https://doi.org/10.1021/acs.est.3c06012>
- Landis, J. D. (2025). “Data for Modern Soil Chronometry using Fallout Radionuclides”, Mendeley Data, V2, doi: 10.17632/cfxkpn6hj9.1.
- Landis, J. D., Hicks Pries, C., McFarlane, K., & Renshaw, C. E. (2025). Direct measurement of soil carbon sequestration over years to centuries. *In Review*, preprint available at <https://doi.org/10.21203/rs.3.rs-5505878/v1>.

- 575 Landis, J. D., Obrist, D., Zhou, J., Renshaw, C. E., McDowell, W. H., Nytch, C. J., Palucis, M. C., Del Vecchio, J., Lopez, F. M., & Taylor, V. F. (2024). Quantifying soil accumulation of atmospheric mercury using fallout radionuclide chronometry. *Nature Communications*, 15, 5430. <https://doi.org/10.1038/s41467-024-49789-7>
- Landis, J. D., Renshaw, C. E., & Kaste, J. M. (2012). Measurement of ⁷Be in soils and sediments by gamma spectroscopy. *Chemical Geology*, 291, 175–185. <https://doi.org/10.1016/j.chemgeo.2011.10.007>
- 580 Landis, J. D., Renshaw, C. E., & Kaste, J. M. (2016). Beryllium-7 and lead-210 chronometry of modern soil processes: The Linked Radionuclide aCcumulation model, LRC. *Geochimica et Cosmochimica Acta*, 180, 109–125. <https://doi.org/10.1016/j.gca.2016.02.013>
- 585 Landis, J. D., Sharma, M., Renock, D., & Niu, D. (2018). *Rapid desorption of radium isotopes from black shale during hydraulic fracturing. 1. Source phases that control the release of Ra from Marcellus Shale.* <https://doi.org/10.1016/j.chemgeo.2018.06.013>
- Landis, J. D., Taylor, V. F., Hintellman, H., & Hrenchuck, L. (2024). Predicting behavior and fate of atmospheric mercury in soils: age-dating METAALICUS Hg isotope spikes with fallout radionuclide chronometry. *Environmental Science and Technology*, in press.
- 590 Mairing, A., & Gáfvert, T. (2013). Radon tightness of different sample sealing methods for gamma spectrometric measurements of ²²⁶Ra. *Applied Radiation and Isotopes*, 81, 92–95. <https://doi.org/10.1016/j.apradiso.2013.03.022>
- 595 Murray, A. S., Helsted, L. M., Autzen, M., Jain, M., & Buylaert, J. P. (2018). Measurement of natural radioactivity: Calibration and performance of a high-resolution gamma spectrometry facility. *Radiation Measurements*, 120, 215–220. <https://doi.org/10.1016/j.radmeas.2018.04.006>
- 600 Murray, A. S., Marten, R., Johnston, A., & Martin, P. (1987). Analysis for naturally occurring radionuclides at environmental concentrations by gamma spectrometry. *Journal of Radioanalytical and Nuclear Chemistry*, 115(2), 263–288.
- 605 Pham, M. K., Sanchez-Cabeza, J. A., & Povinec, P. P. (2005). *Report on the worldwide intercomparison exercise IAEA-385 radionuclides in Irish sea sediment.*
- Pham, M. K., Sanchez-Cabeza, J. A., Povinec, P. P., Andor, K., Arnold, D., Benmansour, M., Bikit, I., Carvalho, F. P., Dimitrova, K., Edrev, Z. H., Engeler, C., Fouche, F. J., Garcia-Orellana, J., Gascó, C., Gastaud, J., Gudelis, A., Hancock, G., Holm, E., Legarda, F., ... Wyse, E. (2008). A new Certified Reference Material for radionuclides in Irish sea sediment (IAEA-385). *Applied Radiation and Isotopes*, 66(11), 1711–1717. <https://doi.org/10.1016/j.apradiso.2007.10.020>
- 610 Sanchez-Cabeza, J. A., & Ruiz-Fernández, A. C. (2012). ²¹⁰Pb sediment radiochronology: An integrated formulation and classification of dating models. *Geochimica et Cosmochimica Acta*, 82, 183–200. <https://doi.org/10.1016/j.gca.2010.12.024>
- 615 Syam, N. S., Lim, S., Lee, H. Y., & Lee, S. H. (2020). Determination of radon leakage from sample container for gamma spectrometry measurement of ²²⁶Ra. *Journal of Environmental Radioactivity*, 220–221. <https://doi.org/10.1016/j.jenvrad.2020.106275>
- 620 Tasker, T. L., Burgos, W. D., Ajemigbitse, M. A., Lauer, N. E., Gusa, A. V., Kuatbek, M., May, D., Landis, J. D., Alessi, D. S., Johnsen, A. M., Kaste, J. M., Headrick, K. L., Wilke, F. D. H., McNeal, M., Engle, M., Jubba, A. M., Vidic, R. D., Vengosh, A., & Warner, N. R. (2019). Accuracy of methods for reporting inorganic element concentrations and radioactivity in oil and gas wastewaters from the Appalachian Basin, U.S. Based on an inter-laboratory comparison. *Environmental Science: Processes and Impacts*, 21(2), 224–241. <https://doi.org/10.1039/c8em00359a>
- 625

Zhang, W., Ungar, K., Chen, J., St-Amant, N., & Tracy, B. L. (2009). An accurate method for the determination of ^{226}Ra activity concentrations in soil. *Journal of Radioanalytical and Nuclear Chemistry*, 280(3), 561–567. <https://doi.org/10.1007/s10967-009-7452-3>